



# Type examination certificate

## No. CH-MI001-09024-01

*Bauartprüfzertifikat*

*Applicant:*  
*Auftraggeber:*

**G2 MISURATORI SRL**  
Via san Martino 38  
14100 Asti  
Italy

*Requirements:*  
*Anforderungen:*

Swiss ordinance on measuring instruments (SR 941.210) of February 15, 2006 annex 2 module B;

Directive 2004/22/EC of the European Parliament and Council of March 31, 2004 on measuring instruments (MID), appendix B, class of measuring instruments MI-001.

Conformity of superseded standard:  
EN 14154 Edition 2007

*Type of instrument:*  
*Geräteart:*

**Single-jet impeller meter intended for the metering of cold and hot water**

*Type designation:*  
*Typenbezeichnung:*

**SFU / SFUP / SFUL  
SCU / SCUP / SCUL**

*Accuracy class(es):*  
*Genauigkeitsklasse(n):*

2 (according to OIML R 49)

*Characteristics:*  
*Kenndaten:*

PN 16, DN 15 / 20, T30, T50, T70, T90, Class B, E1,  
Q<sub>3</sub> 1.6 / 2.5 / 4.0

*Certificate valid until:*  
*Zertifikat gültig bis:*

July 6, 2019

CH-3003 Bern-Wabern, November 10, 2009

*Notified body:*  
*Benannte Stelle:*

Certificate authority METAS-Cert  
No. 1259

*For the test:*  
*Für die Prüfung:*

Dr. Hugo Bissig

Jürg Ramseyer, Head of METAS-Cert